

# Device Screening



## Screening Protocol

Opportunities for patient to identify device:

- Questionnaire
- Technologist review
- Verbal verification



## Device Identification

- Manufacturer, model and serial number
- MR labeling
- Document in medical record



## During Exam

- Closely monitor patient
- Examine localizer images



## Post Exam

- Evaluate patient
- Device prepared or reprogrammed, if required

## GO

- ✓ MR Safe Label

## CAUTION

- ! MR Conditional Label
- ! No Label for MR
- ! Proceed only after all the following are met:
  - › MR system meets device conditions
  - › Device prepared or programmed, if required
  - › Risk-benefit assessment by a radiologist

## STOP

- ✗ MR Unsafe Label
- ✗ MR System Conditions Not Met
- ✗ Device Conditions Not Met
- ✗ Unidentified Device

