

Device Screening



Screening Protocol

Opportunities for patient to identify device:

- Questionnaire
- Technologist review
- Verbal verification



Device Identification

- Manufacturer, model and serial number
- MR labeling
- Document in medical record



During Exam

- Closely monitor patient
- Examine localizer images



Post Exam

- Evaluate patient
- Device prepared or reprogrammed, if required

GO

- ✓ MR Safe Label

CAUTION

- ! MR Conditional Label
- ! No Label for MR
- ! Proceed only after all the following are met:
 - › MR system meets device conditions
 - › Device prepared or programmed, if required
 - › Risk-benefit assessment by a radiologist

STOP

- ✗ MR Unsafe Label
- ✗ MR System Conditions Not Met
- ✗ Device Conditions Not Met
- ✗ Unidentified Device

